	<h1>Certificate of Calibration</h1>	Certificate Serial No:	N/A
		Date of Issue:	12/12/2017
		SATRA Registration No:	SER 2170 SER 2171 SER 2172 SER 2173
		Page 1 of 2	

Wyndham Way, Telford Way, Kettering, Northamptonshire, NN16 8SD, England
 Telephone: +44 (0)1536 410000, Facsimile: +44 (0)1536 410626, e-mail: quality.services@satra.co.uk

Equipment Description: Thermocouples	Equipment Model Number: HI98804	Equipment Serial Number: N/A
Equipment Manufacturer: Hanna	Calibration Procedure Used: HML05	Certificate Serial Number: N/A
SATRA Registration No: SER 2170, 2171, 2172 and 2173	Specification / Test Method: 1±0.2°C	Location of Calibration: N/A

CONCLUSIONS



SER 2170 requires a correction of x 0.9990-1.3844
 SER 2171 requires a correction of x 1.0018-1.5222
 SER 2172 requires a correction of x 0.9987-1.5161
 SER 2173 requires a correction of x 0.9939-1.4019

The uncertainty of measurement for the thermocouples are ± 0.35°C


The calibration results are presented on the following pages.

NOTES

This calibration certificate provides traceability of measurement to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories.

Calibrated By: Signed:	Hannah Pratt 	Authorised By: Signed:	Kaileigh Heels 
Date: 6/12/2017		Date: 12/12/2017	

The reported uncertainties are each based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.

	<h1>Certificate of Calibration</h1>	Certificate Serial No:	N/A
		Date of Issue:	12/12/2017
		SATRA Registration No:	SER 2170 SER 2171 SER 2172 SER 2173
		Page 2 of 2	



MEASURED VALUES

Reference temperature SER 1709	SER 2170	SER 2171	SER 2172	SER 2173
°C				
23.1	24.7	24.7	24.7	24.7
-23.2	-22.0	-21.8	-21.7	-21.8
68.9	70.4	70.4	70.4	70.4
150.2	151.5	151.3	151.8	152.5
200.5	202.2	201.7	202.4	203.3

slope	0.9990	1.0018	0.9987	0.9939
intercept	-1.3844	-1.5222	-1.5161	-1.4019

r ² value	1.0000	1.0000	1.0000	1.0000
----------------------	--------	--------	--------	--------

$$\text{Correction} = \text{Value} \times \text{Slope} + \text{Intercept}$$

Calibrated By:	Hannah Pratt	Authorised By:	Kaileigh Heels
Signed:		Signed:	
Date:	6/12/2017	Date:	12/12/2017

The reported uncertainties are each based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%.