



# CERTIFICATE OF TEMPERATURE MAPPING

ISSUED BY: **Arena Instrumentation Ltd**

Start Time and Date of Mapping: 29<sup>th</sup> January 2019  
End Time and Date of Mapping: 12<sup>th</sup> February 2019

Date of Issue: 15<sup>th</sup> February 2019



**ARENA INSTRUMENTATION LTD**  
Unit 1, Rossmore Business Village,  
Inward Way, Ellesmere Port, Cheshire CH65 3EY  
Tel: 0151 355 1314  
Website: [www.arena-instrumentation.com](http://www.arena-instrumentation.com)  
Email: [info@arena-instrumentation.com](mailto:info@arena-instrumentation.com)

Approved Signatory

*M. Horsley*  
M. Horsley

Customer: PM Strategic Sourcing Ltd  
Unit 1b South Park Court,  
Hobson Street, Macclesfield,  
Cheshire, SK11 8BS

Manufacturer: PM Strategic Sourcing  
Description: Cool Bag  
Model: N/A  
Serial: N/A

Location: Unit 5 Loughborough

Operator: Mark Horsley

Our Ref: 15731

<u>CONDITION OF UNIT UNDER TEST</u>	<u>YES/NO</u>
The unit under test was adjusted	N
The unit under test was repaired	N

## STABILITY

The readings given are the results at the time of mapping and do not carry any implication regarding the long terms stability of the unit under test.

## ACCREDITATIONS

Arena Instrumentation is accredited by UKAS to BS EN 17025:2005 to undertake the calibration presented in this certificate.

## ENVIRONMENT

The instrument was mapped on site with the ambient conditions between 22°C, 25°C and 32°C.

## PROCEDURE

AINS1.008

## UNCERTAINTIES

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a coverage probability of approximately 95 %. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

**NOTES**

1. The Unit Under Test was switched on for a minimum of 60 minutes prior to mapping.
2. The temperatures quoted are with respect to the international temperature scale, ITS-90.
3. This mapping applies only to Unit Under Test.
4. TESTING FOR TEMPERATURE STABILITY OF 2-8 DEGREES (4°C Used)
  - Pre freeze the boards for 12 hours
  - Place 2 frozen boards in the base of the transport case and 2 frozen boards in the lid of the case in the mesh insert
  - Cover the ones in the base with a layer of bubble wrap
  - After a period of half an hour (to allow the case to cool) fill the case with goods (for the purpose of this test ensure the case is full of bubble wrap)
  - Close the case
  - Commence test
5. TESTING FOR TEMPERATURE STABILITY OF 15-25 DEGREES (20°C Used)
  - Pre freeze the freeze boards for 12 hours
  - Place 2 frozen boards in the base of the transport case and 2 frozen boards in the lid of the case in the mesh insert
  - After a period of half an hour (to allow the case to cool) remove all freeze boards
  - Fill the case with goods (for the purpose of this test ensure the case is full with bubble wrap)
  - Close the case
  - Commence test

## Mapping Report

This document has been reviewed and approved POST execution by the undersigned

Approved By (Print)	Job Title:	Signature:	Date:
Mark Horsley	Arena Technician	Mark Horsley	15 <sup>th</sup> February 2019
	Equipment Owner		
	Department Manager		
	QA/Ops Manager		

**Summary**

Issues	User Defined	Signature:	Date:
Hot Spots	User Defined	Mark Horsley	15 <sup>th</sup> February 2019
Cold Spots	User Defined	Mark Horsley	15 <sup>th</sup> February 2019
Other Issues	User Defined	Mark Horsley	15 <sup>th</sup> February 2019

### 1.0 Scope

Temperature Limits	Humidity Limits	Logging Period	Logging Interval
None Standard (User defined LIMITS)	N/A	24 Hours	5 minutes
Core/Load Probe Used	Y		

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

## 2.0 Method of Calibration

The unit under test was mapped in accordance with work instruction AINS1.008 Mapping of Equipment and Rooms.


The expanded uncertainty of individual measurements is

$\pm 0.21\text{ }^{\circ}\text{C}$

*The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a coverage probability of approximately 95 %.*

## 3.0 Layout

The diagram shown below indicates position of the data loggers.



## 4.0 Loading





This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.



This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.


## 5.0 Results

### Temperature against Time Plot




This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

723 Amb 22c, 724/794 Log 20c



723 Amb 25c, 724/794 Log 20c



This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

## 6.0 Test Equipment

Ref	Equipment Type	Manufacturer / ID Number	Model	Serial Number	Calibration Certificate Number	Calibration Date
AI-L44	Data Logger	Gemini / Tiny Tag	TK-4014/TK-4023	712-724 & 794	AI-L44/2018	31 <sup>st</sup> July 2018

## 7.0 Summary

Temperature excursions outside the temperature limits are to be determined by the user.